

 <p>IIT PALAKKAD</p>	<p>भारतीय प्रौद्योगिकी संस्थान पालक्काड Indian Institute of Technology Palakkad अहलिआ एकीकृत कैम्पस, कोज़िपारा Ahalia Integrated Campus, Kozhipara पालक्काड- 678557 Palakkad – 678 557</p>	<p>दूरभाषसंख्या/ Phone no: 04923 – 226300/590/586 ईमेल/ Email : purchase@iitpkd.ac.in</p>
-------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------

Ref No: IITPKD/ELE/AA/106/2017

Dt: 23.01.2018

CORRIGENDUM

Sub: Corrigendum- reg.,

Ref No: Tender no: IITPKD/ELE/AA/106/2017, Dated: 18.01.2018,

Due Date: 02.02.2018 AT 3.00PM

SI No	Existing Entry	To be read as
1.	Part A Benchtop 6 ½ Digit Multimeter 4 (C) DC Current Measurement Accuracy- At least 60 ppm	Part A Benchtop 6 ½ Digit Multimeter 4 (C) Accuracy- At least 600 ppm
2.	6 (D) Resistance Measurement Accuracy - At least 40 ppm	6 (D) Resistance Measurement Accuracy - At least 200 ppm
3.	7 (a) Capacitance Measurement Range -At least upto 100 uF	7 (a) Capacitance Measurement Range -At least upto 100 mF

Rest of the terms and conditions of IIT Palakkad will remain unchanged.

UKJ

REGISTRAR, IIT PALAKKAD