Indian Institute of Technology Palakkad
भारतीय प्रौद्योगिकी संस्थान पालक्काड
STORES & PURCHASE SECTION
Email: purchase@iitpkd.ac.in
Telephone: +91-491 209 2062/63
GSTIN: 32AAAAI9910J1ZR

Tender No. TENDER/2026-27/357
Date of Publication: 25-11-2026
Date/Time of Closing: 16-12-2026, 15:00 hours

Indian Institute of Technology Palakkad Invites Tender under Two-bid system for the:

SUPPLY, INSTALLATION, TESTING AND COMMISSIONING OF ACCELERATED WEATHERING TESTER

Conforming to the specifications as in BoQ Technical.

Tender Documents may be downloaded from the e-Wizard Portal https://mhrd.ewizard.in/. Aspiring Bidders who have not enrolled / registered in e-Wizard should enroll / register before participating through the website https://mhrd.ewizard.in/. Bidders are advised to go through instructions provided at **"Procedure for Submission of E-tender".** [Special Instructions to the Contractors/Bidders for the e-submission of the bids online through this e-Wizard Portal"].
Bidders can access tender documents on the website. For searching in the site, kindly go to Live Tenders option, Click "Advance Search" and select the Department as 'IIT Palakkad'. Thereafter, Click on "Search" button to view all IIT Palakkad tenders. Select the appropriate tender and fill them with all relevant information and submit the completed tender document online on the website https://mhrd.ewizard.in/ as per the timeline below.

No manual bids will be accepted. All tender documents including Techno-Commercial, Technical and Financial bids should be submitted in the e-Wizard portal.

S. No.	Events	Date and Time
1	**Publication of the Tender Document**	**25-11-2026**
2	**Last Date/Time for submission of ONLINE Bids**	**16-12-2026, 15:00 hours**
3	**Opening of Technical Bids**	**16-12-2026, 15:15 hours**

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

TERMS AND CONDITIONS

1	**GENERAL**	a) The responsibility of submission of the bids on or before the last date shall rest with the tenderer. The institute will hold no responsibility for the non-receipt of the bids or the bids received after the date/time specified. Any bid received by IITPKD after the bid submission deadline prescribed by IITPKD, shall be rejected and returned unopened to the Bidder. b) Canvassing or offer of an advantage or any other inducement by any person with a view to influencing acceptance of a bid is an offence under the Laws of India. Such action will result in the rejection of bid, in addition to other punitive measures. c) Each bidder shall submit only one bid, either by himself or as a partner in a joint venture or as a member of the consortium. If a bidder or if any of the partners in a joint venture or any one of the members of the consortium participate in more than one bid, the bids (of both the individual and the partnership/consortium/joint venture) are liable to be rejected. d) The bidder shall bear all costs associated with the preparation and submission of his bid and IITPKD shall in no case be responsible or liable for those costs, regardless of the conduct or outcome of the tender process. e) **IITPKD will respond to any request for clarification or modification of the Tender Document that are received up to TWO DAYS prior to the deadline for submission of bids prescribed by IITPKD. For this purpose, the prospective bidder(s) requiring clarification in the Tender Document shall notify IITPKD through the ONLINE Portal ONLY. Any such clarification, together with all the details on which the clarification had been sought, will be published in the ONLINE Portal ONLY. Deviations, if any, observed by the Institute in the submitted bids, from the Terms and Conditions of the Tender Document will not be accepted by the Institute.** f) Except for any such clarification by the Institute, which is expressly stated to be an addendum to the tender document issued by the Registrar, IIT Palakkad, no written or oral communication, presentation or explanation by any other employee of any of the Sections/Departments of the Institute, shall be taken to bind or fetter the Institute. g) The bidder is expected to examine all instructions, forms, terms and conditions in the Tender Document. In the event of discovery of any missing pages, the bidder shall inform the same to the Section/ Department concerned. Failure to furnish the information required by the Tender Document or submission of a tender not substantially responsive to the Tender Document in every respect will be at the bidder's risk and may result in rejection of the bid. h) The bidder shall not make or cause to be made any alteration, erasure or obliteration to the text of the Tender Document. i) The Supplier shall not, without the prior written consent of the IITPKD, assign to any third party, the Contract or any part thereof.
2	**COMPOSITION OF THE TENDER DOCUMENT**	(a) The Tender Document comprises of: Instruction to the bidders including terms and conditions 1. Technical Specifications (Annexure-I) 2. Undertaking by the Bidder (Annexure-II) 3. Fall Clause Notice Certificate (Annexure-III) 4. Annexure regarding Blacklisting/Debarment (Annexure-IV)

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

3	**DOCUMENTS COMPRISING THE BID**	• **The Technical, Techno-commercial (Cover One), and Commercial Bid (Cover Two) shall be submitted ONLINE through the e-Wizard Portal.** • Bids submitted in any mode other than ONLINE will be rejected outright. • Documents establishing the conformity of the terms and conditions of the Tender Document shall be provided along with the bid. The offer/bids should be sent only for a system or that is available in the market and supplied to a number of customers. A list of customers in India and abroad with details must accompany the quotations. Quotations for a prototype machine will not be accepted. • Original catalogue (not any photocopy) of the quoted model duly signed by the principals must accompany the quotation in the Technical bid. No prices should ever be included in the Technical bid. • Compliance or Confirmation report with reference to the specifications and other terms and conditions should also be obtained from the principal. • Information related to the agency/bidder such as photocopies of the Registration/PAN/GST/TIN shall be furnished. • The technical bid should consist of all technical details along with commercial terms and conditions. **No prices should be included in the technical bid. Mentioning of Prices in the Technical Bid shall lead to DISQUALIFICATION.** • Bidders/OEM who are bidding for this tender, a) Should have implemented at least **ONE ORDER** for **Accelerated Weathering Tester during the previous four financial years (2022-23, 2023-24, 2024-25 and 2025-26)** from Centrally Funded Technical Institutes (IITs, NITs, IISc, IISER), DRDO, ISRO, CSIR labs or Government Firms in India. Copies of the most recent purchase orders and user certificates of successful implementation must be included. Copies of financial statements or evidence of turnover must be furnished. b) Should have an **Average Annual Turnover of Rs.65,00,000/- (RUPEES SIXTY-FIVE LAKH ONLY) during the last THREE financial years (2023-24, 2024-25, and 2025-26).** The bidder shall enclose the audited statements of the indicated financial years, which should have been certified by a Chartered Accountant or a Competent Authority. c) Should submit a Digitally signed Tender Document in Cover One.
4	**EARNEST MONEY DEPOSIT (EMD)**	• The bidder shall furnish EMD of **Rs.64,000/- (Rupees Sixty-Four Thousand Only)** through online payment gateway in the E-Wizard. • **Bids not accompanied by EMD shall be DISQUALIFIED.** • The firms who are registered with National Small Industries Corporation (NSIC) / or Small Scale Industrial (SSI)/ Micro & Small Enterprises (MSEs) are exempted from submitting the EMD. NSIC / MSME registered bidders must submit a copy of a valid NSIC / MSME Registration Certificate for exemption of EMD. It will be applicable for those bidders who shall produce their own goods or provide their own services, and not applicable for trading purposes.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

5	**PERFORMANCE SECURITY**	● The performance security shall be submitted within **FIFTEEN DAYS** of receipt of the material by the IITPKD. The successful bidder shall furnish the Performance Security equal to **5%** of the order / contract value (excluding the value of annual maintenance charges). The Performance Security shall be valid all along the warranty period and shall extend up to **SIXTY DAYS** after the date of completion of warranty period. It shall be ensured by the successful bidder that the validity of the Performance Security submitted is extended depending on the date of commencement of the Warranty. ● The performance security shall be a bank guarantee/E-Bank Guarantee (in the format as provided) issued by the Indian Scheduled bank acceptable to the IITPKD or a Demand Draft favoring, INDIAN INSTITUTE OF TECHNOLOGY PALAKKAD payable at PALAKKAD. ● The performance security shall automatically become null and void once all the obligations of the Supplier under the Contract have been fulfilled, including, but not limited to, any obligations during the Warranty Period and any extensions to the period. The performance security shall be returned to the Supplier not later than fifteen (15) days after its expiration. ● Failure of the successful Bidder to comply with the requirements shall constitute enough grounds for the annulment of the award and forfeiture of the EMD, in which event the IITPKD may make the award to the next lowest evaluated bid submitted by a qualified Bidder or call for new bids.
6	**BID PRICES AND CURRENCY**	● Prices must be quoted separately for each equipment/item identified. ● **Price quoted for equipment/items shall include all the costs associated with packing, local transportation from the point of clearance to IITPKD, insurance, loading, unloading and associated delivery charges. The delivery shall be on DOOR DELIVERY basis to the institute including its installation, commissioning, integration and validation. It is the sole responsibility of the supplier to ensure that the equipment is delivered on DDP mode to IIT Palakkad. An undertaking to this effect as in Annexure-II**. ● Prices quoted by the bidder shall be fixed during the validity of the bid. ● **Prices of the equipment/items shall be quoted in Indian Rupees (INR) / Foreign Currency.**
7	**LETTER OF CREDIT**	● Upon the successful Bidder's furnishing of the copy of the Purchase Order duly signed on each page and the Performance Security, for the equipment ordered in foreign currency, IITPKD will open a letter of credit (LC) in a convenient Nationalized Bank in India. For opening of LC necessary information shall be provided by the supplier or its authorized agents. ● In case the successful bidder is a foreign company and wishes to submit Performance Security in the form of Bank Guarantee, the Bank Guarantee should be routed through the Beneficiary Bank to the end user bank. Otherwise, the Indian Agent of the foreign vendor shall submit a Bank Guarantee from a Nationalized Bank of India. The following documents shall be submitted in case of an Indian agent submitting the Performance Security on behalf of his principal: ● Foreign principal's proforma invoice indicating the commission payable to the Indian agent and nature of after-sales service to be rendered by the Indian agent. ● Copy of the agency agreement with the foreign principal and the precise relationship between them and their mutual interest in the business.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

		● For imported equipment, a Letter of Credit (LC) shall be opened for 100% CIP price on receipt of the acknowledgment of the purchase order. However, 80% of the LC amount only shall be released on proof of the shipment of the consignment with necessary documents to be provided in detail at the time of placing of the purchase order. Balance 20% of the LC amount shall be released upon the receipt of a performance security of 5% of the total value of the purchase order and installation, commissioning, integration, validation and installation report/certification jointly given by the end user and the supplier. **Any costs associated with the amendments made in the LC as per the request made by the Supplier s should be borne by the supplier.** **Any fluctuation in rates / rate conversions arising due to the amendment requests made by the supplier shall be on the supplier and not on the institute.**
8	**PAYMENT TERMS**	100% payment shall be made upon successful supply, installation, testing and commissioning of the equipment, certification by the End User, and submission of the 5% Performance Bank Guarantee (PBG).
9	**PERIOD OF VALIDITY OF BIDS**	● Bids shall remain valid for a period of **180 DAYS** after the date of the deadline for submission of bids prescribed by IITPKD. ● If the deadline is extended due to unforeseen circumstances, the bid validity shall be deemed to have extended accordingly.
10	**TIME FOR SUPPLY, INSTALLATION, COMMISSIONING AND VALIDATION OF THE EQUIPMENTS/ITEMS**	a. The Supplier shall supply the equipment/items within the period specified in the tender document i.e. within **16 WEEKS** of signing the purchase order or within the period mutually agreed between IITPKD and supplier. All the equipment and accessories should be delivered at **IIT Palakkad, Lab-5, C06 Building, Sahyadri Campus, Kanjikode West, Palakkad, Kerala - 678623.** b. The Supplier shall thereafter proceed with the installation, commissioning, integration and validation and demonstrate operational acceptance of the equipment/items within the period specified. The equipment/items shall be installed and commissioned by the successful bidder within 20 to 25 days from the date of its receipt. c. The tenderer should indicate clearly the time required for delivery of the item. In case there is any deviation in the delivery schedule, liquidated damages clause will be enforced or penalty for the delayed supply period will be levied. d. In the event of failure of supply of the item/equipment/items within the stipulated delivery schedule, IITPKD has all the right to purchase the item/equipment/items from other sources on the total risk of the Supplier under the risk purchase clause.
11	**PRODUCT UPGRADES**	The Supplier shall continue to support and maintain the version/model of the Equipment supplied by upgrading the software and the hardware as and when amendments are carried out in the existing version or the product is upgraded. Whereas upgrades to the software shall be supplied free of cost, the Supplier may charge for upgrade in hardware provided it is of major nature. An upgraded higher version of the instrument and software related with the instrument shall be supplied.
12	**PENALTIES**	If the Supplier fails to complete any of the activities in accordance with the time specified for it, or any extension of time granted by IITPKD, Liquidated Damages Clause shall be invoked.
13	**UP-TIME GUARANTEE/ DOWNTIME PENALTY CLAUSE**	The Supplier should provide up-time guarantee of 95% [24 (hours) X 7 (days) X 365 (days)] basis during the warranty period. The Supplier should provide up-time guarantee of 95% (24 hours/day basis) both during warranty. If downtime exceeds the 5% limit, extension of the warranty period will be twice the excess down time period.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

14	**LIQUIDATED DAMAGES**	If a firm accepts an order and fails to execute the order, in full or part, as per the terms and conditions stipulated therein, it will be open to the Institute to recover liquidated damages from the firm at the rate of 1% of the value of the undelivered goods per month or part thereof, subject to a maximum of 5% of the value of the undelivered goods. It will also be open to the Institute alternatively, to arrange procurement of the required stores from any source, at the risk and expense of the firm, accepted and failed to execute the order according to stipulations agreed upon. This will also entail the removal of the defaulters' name from the approved/registered list of Suppliers.
15	**EFFECT OF FORCE MAJEURE**	(a) If the Supplier is prevented, hindered, or delayed from or in performing any of its obligations under the Contract by an event of Force Majeure, then it shall notify the IITPKD in writing of the occurrence of such event and the circumstances of the event of Force Majeure within **FIFTEEN DAYS** after the occurrence of such event. (b) The Supplier, when affected by the event of Force Majeure shall use reasonable efforts to mitigate the effect of the event of Force Majeure upon its performance of the Contract and to fulfill its obligations under the Contract, but without prejudice to IITPKD's right to terminate the Contract. (c) No delay or non-performance by the Supplier caused by the occurrence of any event of Force Majeure shall: ● Constitute a default or breach of the Contract; ● Give rise to any claim for damages or additional cost or expense occasioned by the delay or non-performance. (d) If the performance of the Contract is substantially prevented, hindered, or delayed for a single period of more than **THIRTY DAYS** or an aggregate period of more than **SIXTY DAYS** on account of one or more events of Force Majeure, the IITPKD shall have the right to terminate the Contract by giving a notice to the Supplier.
16	**EXTENSION OF TIME LIMITS FOR SUPPLY AND MAKING OPERATIONAL, THE EQUIPMENT**	● The time limit for supply, installation & commissioning, integration & validation shall be extended if the supply is delayed or impeded in the performance of any of its obligations under the Contract by reason of any of the following: ● Any occurrence of Force Majeure; ● Any other matter specifically mentioned in the Contract; ● By such period as shall be fair and reasonable in all the circumstances and as shall fairly reflect the delay or impediment sustained by the Supplier.
17	**GOVERNING LAW AND SETTLEMENT OF DISPUTES**	(a) The Contract shall be governed by and interpreted in accordance with the laws of India. (b) Any dispute or claim arising out of/relating to this Contract of the breach, termination or the invalidity thereof, shall be settled by the Hon'ble Courts of Justice at Palakkad. **(c)** The page number should be marked in all pages serially (including all supporting documents enclosed with the tender document) and the declaration for the same shall be submitted by the bidder as in **Annexure-II.** (d) IITPKD reserves the right to accept or reject any or all the tenders in part or whole or may cancel the tender at its sole discretion without assigning any reason whatsoever. No further correspondence in this regard will be entertained.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

AWARD OF CONTRACT

1	**AWARD CRITERIA**	1. The Contract will be awarded to the bidder whose offer is technically compliant in all respects and whose financial bid is the lowest among all technically qualified bidders. 2. The Institute reserves the right to buy different items/quantities from different bidders considering price of individual/group of equipment/items or any other factors as decided by the Committee.
2	**AWARD OF PURCHASE ORDER**	a. Prior to the expiration of the period of bid validity, IITPKD will issue the Letter of Intent / Purchase Order to the successful Bidder in writing. b. Any amendment(s) in the Purchase Order will be permitted within **SEVEN DAYS** of its issuance. No amendments will be permitted beyond this period. c. The Purchase Order will constitute the foundation of the Contract.
3	**CONTRACT AGREEMENT**	a. Within **SEVEN DAYS** of receipt of the Purchase Order, the successful Bidder shall sign and date its copy on each page and return it to the Purchaser. b. Copy of Purchase Order duly signed and dated by the successful Bidder on each page shall constitute the Contract Agreement.
4	**CONTRACT DOCUMENTS / AMENDMENT TO CONTRACT**	a. All documents forming part of the Contract (and all parts of these documents) are intended to be correlative, complementary and mutually explanatory. The Contract shall be read as a whole. b. The order of precedence of the Contract documents shall be as follows: i. Contract Agreement/Purchase Order ii. All Forms/Annexures iii. equipment/items and their requirement iv. Supplier's Bid v. Tender Document c. No amendment or other variation of the Contract shall be effective unless it is in writing, is dated, expressly refers to the Contract and is signed by a duly authorized representative of each party to the Contract.

REGISTRAR

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

Name of the Item :	ACCELERATED WEATHERING TESTER
Quantity:	1
Warranty Period:	3

TECHNICAL SPECIFICATIONS

S. No.	Items	IIT PKD required Specification
1	**Equipment Type and Relevant Standards**	The equipment shall be an accelerated fluorescent UV accelerated weathering instrument with condensation and direct water spray, capable of exposing specimens to controlled UV radiation, temperature and moisture in accordance with ASTM G154, ASTM C1442, ASTM D4799, and ISO 4892-3 standards.
2	**General Features**	● The equipment shall be a bench-top/floor-standing, cabinet-type accelerated weathering tester suitable for laboratory use and for continuous, unattended, programmed operation. ● Should be suitable for exposing construction materials such as bituminous binders, asphalt mixtures, polymers, rubbers, coatings, cementitious materials and geo-materials for controlled the UV radiation intensity, moisture and temperature. ● The offered configuration shall support ASTM G154, ASTM C1442, ASTM D4799, ISO 4892-3 and other relevant standard test methods as specified by the purchaser.
3	**Technical Requirements**	● The UV source shall consist of eight fluorescent UV lamps arranged to provide uniform specimen exposure. ● The equipment shall support UVA-340, UVA-351, UVB-313 or other compatible lamps as required by the applicable test method. A complete set of UVA-340 lamps shall be supplied as standard. ● The irradiance shall be controlled automatically through a closed-loop system using a UV sensor and feedback mechanism that continuously monitors the irradiance and adjusts the lamp power to maintain the programmed set point. Manual control systems are not acceptable.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

		● The system shall automatically compensate for lamp ageing and other variability, and shall operate the lamps until the set point can no longer be maintained, so as to ensure reproducibility and repeatability of results. ● The irradiance set point shall be adjustable over a range of at least 0.25 – 1.55 W/m²/nm at 340 nm (UVA-340 lamps) or better, with a control accuracy suitable for ASTM G154 and ISO 4892-3. ● The irradiance sensor shall be a wideband UV sensor covering approximately 250–400 nm or wider. ● An automatic, onboard irradiance calibration facility shall be provided, using a compatible calibrated radiometer/sensor, such that no separate calibration device or manual intervention is required for routine calibration. Calibration shall be traceable to national/international standards. ● Programmable UV exposure cycles with adjustable duration shall be provided. ● An integrated condensation system shall be provided to simulate dew and moisture exposure. ● An integrated direct water-spray system shall be provided to simulate rain, thermal shock and erosion. ● Programmable combined cycles of UV, condensation and water spray shall be supported, including the standard cycles of ASTM G154 and ISO 4892-3, and user-defined multi-step cycles. ● Black panel temperature shall be controlled over a range of at least 35 – 80 °C (or better) as appropriate to the test method, with a control tolerance of ±3 °C or better. ● The condensation temperature shall be controlled over a range suitable for the test standards with a tolerance of ±3 °C or better. ● The chamber shall accommodate a minimum of 48 specimens of 75 x 150 mm, providing a total exposure area of approximately 4,000 cm². ● Standard specimen holders for flat panels (up to approximately 20 mm thick) shall be provided.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

		● Special holders shall be provided for large 3D specimens such as asphalt plugs, cores and concrete blocks of circular, semi-circular, and rectangular cross sections with thickness ranging from 50 mm to 200 mm, with the exposed surface kept in the specimen plane so that the irradiance at the specimen surface stays within the set point and the tolerances prescribed in ASTM G154 and ISO 4892-3. ● Water quality requirements shall conform to ASTM G154/ISO 4892-1. A suitable water purification unit shall be quoted as an option. ● Controller and Software Capabilities: o Dual touch-screen displays for real-time viewing of test parameters and performance, and for control of tester operation and settings o Multi-colour LED status indicator showing the operational status of the tester at a glance. o Programmable test-cycle settings, automatic recording of test and performance data (irradiance, temperature, elapsed time and cycle status). o Data transfer through USB and built-in Ethernet connection for data logging. o Built-in self-diagnostic warnings and service reminders. o Suitable software shall be provided for graphical display of results and data export (Excel/CSV/PDF formats) without additional cost. o Free upgrade of software version when available in the future.
4	**Accessories**	● Complete set of specimen holders, and blank panels for filling unused positions, shall be provided. ● A calibrated radiometer/irradiance sensor with calibration certificate, and a black panel (and, where applicable, black standard) thermometer shall be provided. ● One additional set of UV lamps shall be supplied as spares. ● All necessary water-supply and drain connections, tubing, connectors, tools, operating manuals and data-transfer accessories must be provided.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

		● The equipment shall be supplied with a computer and dedicated software for data acquisition, analysis and reporting. ● Standard caster wheels shall be provided for positioning and movement within the laboratory.
5	**Power & Safety**	● Power supply requirement should be suitable for Indian conditions (230 V ±10%, 50 Hz), and the electrical load shall be stated in the offer. ● The equipment shall be of durable, corrosion-resistant construction with a protective enclosure, UV-shielded door and door interlock. ● Over-temperature, low-water-level and lamp-failure protection with audible/visual alarms shall be provided. ● Automatic shutdown or protective mechanisms to prevent equipment damage during sudden power failures, voltage fluctuations or system overloads with a power backup for at least 15 minutes. The test shall resume automatically after power restoration.
6	**Warranty & Support**	● The equipment shall have a minimum 3-year comprehensive warranty. ● Extended warranty beyond 3 years is desirable. ● Installation and commissioning shall be carried out at IIT Palakkad and shall include equipment operation, calibration, data interpretation and routine maintenance. ● The equipment shall be accepted after successful installation and functional demonstration.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST

ANNEXURE - II

UNDERTAKING BY THE BIDDER
(TO BE SUBMITTED ONLY THROUGH ONLINE MODE IN APPROPRIATE FORMAT)

We here by accept all the Terms and Conditions of the Tender Document and strictly adhere to the same in the event of getting Purchase order. We also declare that the Technical and Financial Bids submitted by us has NO DEIVATION from the Tender Terms and Conditions.

We hereby accept that the PRICES OF THE EQUIPMENTS/ITEMS QUOTED ARE AS PER THE INCOTERMS 2022 - DDP MODE, IIT PALAKKAD AND CLAUSE NO.6 OF THE TENDER DOCUMENT.

We hereby undertake that there are _____ pages, serially numbered, in the submitted tender including the supporting documents. (Please serially number all the pages including blank page, if any).

We have submitted our principal's exclusive authorization letter which is specific for this tender No._____ dated _____.

Note: This letter should be on the letterhead of the quoting firm and should be signed by a Competent Authority.

ANNEXURE - III

FALL CLAUSE NOTICE CERTIFICATE
(TO BE SUBMITTED ONLY THROUGH ONLINE MODE IN APPROPRIATE FORMAT)

This is to certify that we have offered the maximum possible discount to you in our Quotation No. _____ dated _____ **(Please do not reveal the prices here, which will lead to outright rejection of your bid).**

The prices charged for the Stores supplied under tender should under no event be higher than the lowest prices at which the party sells the items of identical description to any other Govt. organization/PSU's/Central Govt., /State Govt. Autonomous bodies/Central/state Universities/Central/State Educational Institutions, failing which the "**FALL CLAUSE**" will be applicable. The institute will look into a reasonable past period to ensure this.

In case, if the price charged by our firm is found to be more, **IIT Palakkad** will have the right to recover the excess charged amount from the subsequent/unpaid bill of the supplier.

Note: This letter of authority should be on the letterhead of the quoting firm and should be signed by a Competent Authority and having the power of attorney.

ANNEXURE - IV

UNDERTAKING REGARDING BLACKLISTING / NON – DEBARMENT

Tender No. _____

To,
M/s. Indian Institute of Technology Palakkad
Kanjikode, Palakkad, Kerala 678623.

We hereby confirm and declare that we, M/s _____ are not blacklisted/ De-registered / debarred by any Government department/ Public Sector Undertaking/ Private Sector/ or any other agency for which we have Executed/ Undertaken the works/ Services.

For
Company Name and Seal
Authorised Signatory

Note:This letter should be on the letterhead of the quoting firm and should be signed by a Competent Authority.

Digitally signed by Geetha
Date: 2026.11.25 10:28:25 IST